

# Klargøringsrapport

Duotec A/S

Herstedøstervej 19

2600 Glostrup

Tlf. 4345 9188 - Fax 4343 2330

email: info@duotec.dk

Kunde: PN Engineering			Projekt:			Tekniker: Rasmus Bendsen		
Detektorinformation			Områdejustering			Kalibreringsdata		
Fabrikat	Type	Serienummer	Kalibreret for	Span	Enhed	Testgas	Nulpunkt	Span
RAE	ppbRAE	594-911555	C4H8	1000	ppm	1	0	10

	Gas	Koncentration	Enhed	Tolerance	Certifikat
1	C4H8	10	ppm	2%	1626913-15
2					
3					
4					

29.04.2018

Dato



Underskrift